



ADVANCED FIBER SOLUTIONS

Certificate of Calibration

Advanced Fiber Solutions, Inc
19 Norfolk Ave,
Easton, Ma, 02375, USA
Ph: 1-800-556-9313
Fax: 1-617-507-0784
www.advancedfibersolutions.com

Advanced Fiber Solutions hereby certifies this meter meets or exceeds all published specifications and has been calibrated using standards whose accuracies are traceable to the National Institute of Standards and Technology, within the limitations of the Institutes calibration standards, or have been derived for acceptable values of natural physical constants, or have been derived by the ratio type self-calibrating techniques.

Calibration Standard Used:

Advanced Fiber Solutions Asset# 0002
HP 8153A/HP81533B/HP81525A
Advanced Fiber Solution Calibration Certificate # 1000397525
Calibrated Date: 12/26/2006 Due Date 12/26/2007


Calibrated Equipment:

Model Number: **OM220**
Serial Number: **m711e005**
Calibrated Wavelengths: **850nm, 1300nm, 1310nm & 1550nm**
Calibration Date: **November 2007**
Due Date: **November 2008**

Calibrated By:



Quality Assurance Inspector:



This certificate relates only to the item specified above. This document may not be reproduced without the express written approval from Advanced Fiber Solutions.