

# Optical Source Report Certificate

**Advanced Fiber Solutions, Inc**  
**19 Norfolk Ave,**  
**Easton, Ma, 02375, USA**  
**Ph: 1-800-556-9313**  
**Fax: 1-617-507-0784**  
**www.afs-i.com**

Advanced Fiber Solutions hereby certifies this fiber optic source meets or exceeds all published specifications and has been tested by the below products:

**Products used to create report:**

86140B Agilent Optical Spectrum Analyser

HP 8153A/HP81533B/HP81525A

Advanced Fiber Solution Calibration Certificate # 1000397525

Calibrated Date: 12/26/2007 Due Date 12/26/2008

**Calibrated Equipment:**

Model Number:

**OS417-2MD**

Serial Number:

**S6070014**

Wavelengths:

**665nm & 850nm**

Center Wavelength 850nm LED:

**849.7nm**

Output power:

**850nm > -18.0 dBm into 62.5m fiber**

**665nm > -13.0 dBm into 1000m POF**

Report Date:

**September 2008**

Due Date:

**September 2009**

Report Created By:

